

# Certificate of Quality

## Product Data

Product Description	Conical Centrifuge Tube,PP/HDPE, 50ml, 30x115mm, 25/sleeve, 500/case
Catalogue Number	53050
Lot Number	201/28476
Expiration Date	6 years from manufacture date

## Performed Tests

Test Name	Standard	Results
Cytotoxicity Test	Caused no Cytotoxicity	Complies
Fungistatic Test	EN ISO-11137-2/2012	Complies
Bacteriostatic Test	EN ISO-11137-2/2012	Complies
SAL10 <sup>-6</sup>	EN ISO-11137-2/2012	Complies
Sterility Test	EN ISO-11137-2/2012	Complies
Print Test	Chloroform Proof	Complies
Endotoxin/Pyrogen Test	Gel clot assay ≤0.06 EU/ml	Complies
RNase	Free of any detectable limits	Complies
DNase	Free of any detectable limits	Complies
Graduation test (Maximum Error Limit)	± 0.5ml	Complies
RCF Test	15000 fixed bucket rotor with Methanol	Complies
Autoclave Test	121°C at 15psi for 20 minutes	Complies
Leech Test	As per Bioswisstec AG Qualit Policy	Complies
Leak Test	635mm/Hg or 95 KPA	No leakage
Dimensional	As per Bioswisstec AG Qualit Policy	Complies
Aesthetical defects	No visual defects with 20X Mega Vision	No Defects

Bioswisstec, Ltd. hereby certifies that the product identified above has been inspected in compliance with product quality specification and requirements as documented in our ISO 9001:2008 Quality Management System (K-QA-Q031478) in South-Korea.

This document is valid by electronic signature.

Date (D/M/Y) Quality Control Manager Mr. D. Heger 12.06.2019

For further technical information, please visit : <http://bioswisstec.com>